

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: Cleveland Police

Serial Number: LW06261

Certificate Number: 5243

Date of Calibration: 02 August 2021

Next Calibration: 01 August 2022

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 20.000102 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.3 ft	284.5 ft	488.8 ft

Speed ± 1 MPH

Actual	72 MPH	-35 MPH	-48 MPH	60 MPH	22 MPH
Indicated	72 MPH	-35 MPH	-48 MPH	60 MPH	22 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L Korcz
Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel: 020 8847 4400

email: info@truelouk.com