CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:	Cleveland Police	**Serial Number:**	LW06291
		Certificate Number:	3847
Date of Calibration:	08 February 2020	**Next Calibration:**	07 February 2021

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum		- CC85
Oscillator Frequency		- 19.999901 MHz
Self Test Sequence		- Pass
External Controls		- Pass
Aiming	Vertical	- Pass
	Horizontal	- Pass
Eye Safety		- Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.1 ft	284.5 ft	488.6 ft

Speed ± 1 MPH

Actual	24 MPH	35 MPH	-70 MPH	-60 MPH	54 MPH
Indicated	24 MPH	35 MPH	-70 MPH	-60 MPH	54 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz
Technical Signatory

TRUVELO (UK) LTD

Tel:020 8847 4400 email:info@truvelouk.com