

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: Cleveland Police **Serial Number:** LW06292

Certificate Number: T5117

Date of Calibration: 31 January 2018 **Next Calibration:** 30 January 2019

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999986 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.4 ft	284.6 ft	488.6 ft

Speed ± 1 MPH

Actual	73 MPH	-74 MPH	57 MPH	-25 MPH	31 MPH
Indicated	73 MPH	-74 MPH	57 MPH	-25 MPH	31 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L. Korcz
Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel: 020 8847 4400

email: info@truelouk.com

