

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer: Durham Police

Serial Number: LW06291

Certificate Number: T5254

Date of Calibration: 29 March 2018

Next Calibration: 28 March 2019

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999878 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range \pm 0.5 ft

| | | | |
|-----------|----------|----------|----------|
| Actual | 156.3 ft | 284.5 ft | 488.4 ft |
| Indicated | 156.2 ft | 284.4 ft | 488.6 ft |

Speed \pm 1 MPH

| | | | | | |
|-----------|--------|---------|--------|---------|--------|
| Actual | 73 MPH | -71 MPH | 58 MPH | -26 MPH | 31 MPH |
| Indicated | 73 MPH | -71 MPH | 58 MPH | -26 MPH | 31 MPH |

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK



A Young
Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel: 020 8847 4400

email: info@truvelouk.com